



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQRXXAZ.02

Glu-Plasminogen ELISA Kit 96T

REF TC 12040

LOT 120401B4.01

⌚ 28 - Feb - 2015

| Inhalt Content | Reagenz Reagent | LOT | ⌚ |
|-------------------|--------------------------|---------|-------------|
| 1x | anti Plasminogen Plate | RGX11B2 | 28-Feb-2015 |
| 1x0.3mL | POX anti Plasminogen | RBG21B1 | 28-Feb-2016 |
| 1x80mL | Wash buffer Concentrate | RNA34B0 | 30-Sep-2015 |
| 1x90mL | Inkubation buffer | RNB32B0 | 30-Apr-2015 |
| 1x12mL | TMB Substrat | RKN33B0 | 30-Sep-2015 |
| 1x12mL | Stop solution | RKK33B0 | 31-Oct-2015 |
| 1x1mL | Standard Glu-Plasminogen | RBJ5105 | 31-Jan-2016 |

| Analytische Prozedur Analytical Procedure | | Anforderung Requirement | Ergebnis Result | |
|--|--------------------|----------------------------|--------------------|----|
| OD | Calibrator 1 | > 1 | 1.19 | OK |
| | Calibrator 5 | < 0.2 | 0.05 | OK |
| Wiederfindung der Kontrollen/ Recovery of controls | Internal control 1 | 60 – 250 µg/mL | 68 | OK |
| | Internal control 2 | 60 – 250 µg/mL | 90 | OK |

Durchführung von
Prepared by

Fikret NASUEI
Fikret Nasuei

04-Mär-2014
Datum / Date:

Freigabe von
Approved by

Michael RIHA
Michael Riha

04-Mär-2014
Datum / Date: