

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQRXXAZ.02
---	---	------------

TECHNOZYM® Glu-Plasminogen ELISA Kit 96T

REF TC12040
LOT 120401B5.01
 31-Jan-2016

Inhalt Content	Reagenz Reagent	LOT	
1x	anti Plasminogen Plate	RGX11B3	29-Feb-2016
1x0.3mL	POX anti Plasminogen	RBG21B1	28-Feb-2016
1x80mL	Wash buffer Concentrate	RNA42B0	31-Aug-2016
1x90mL	Inkubation buffer	RNB42B0	30-Apr-2016
1x12mL	TMB Substrat	RKN41B0	31-Aug-2016
1x12mL	Stop solution	RKK43B0	31-Dec-2016
1x1mL	Standard Glu-Plasminogen	RBJ5105	31-Jan-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1	1.17	OK
	Calibrator 5	< 0.2	0.02	OK
Wiederfindung der Kontrollen/ Recovery of controls	Internal control 1	60 – 250 µg/mL	73.5	OK
	Internal control 2	60 – 250 µg/mL	90.5	OK

Durchführung von
Prepared by


Michael RIHA



24-MAR-2015
Datum / Date:

Freigabe von
Approved by

Fikret NASUFI



24-MAR-2015
Datum / Date: