



**ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS**

DQRXXAZ.02

**Glu-Plasminogen ELISA Kit 96T**

**REF** TC 12040

**LOT** 120402B4.01

30 - Nov - 2015

Inhalt Content	Reagenz Reagent	<b>LOT</b>	
1x	anti Plasminogen Plate	RGX11B3	29-Feb-2016
1x0.3mL	POX anti Plasminogen	RBG21B1	28-Feb-2016
1x80mL	Wash buffer Concentrate	RNA35B0	31-Dec-2015
1x90mL	Inkubation buffer	RNB41B0	31-Jan-2016
1x12mL	TMB Substrat	RKN34B0	30-Nov-2015
1x12mL	Stop solution	RKK41B0	31-Mar-2016
1x1mL	Standard Glu-Plasminogen	RBJ5105	31-Jan-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1	1.28	OK
	Calibrator 5	< 0.2	0.01	OK
Wiederfindung der Kontrollen/ Recovery of controls	Internal control 1	60 – 250 µg/mL	139	OK
	Internal control 2	60 – 250 µg/mL	155	OK

Durchführung von  
Prepared by

*Filbert Naselli*

20-Mai-2014  
Datum / Date:

Freigabe von  
Approved by

Michael RIHA

20-MAY-2014  
Datum / Date: