





**ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS**

DQRXXAZ.02


Glu-Plasminogen ELISA Kit 96T

REF TC 12040
LOT 120402B5.01
 28-Feb-2016

Inhalt Content	Reagenz Reagent	LOT	
1x	anti Plasminogen Plate	RGX11B3	29-Feb-2016
1x0.3mL	POX anti Plasminogen	RBG21B1	28-Feb-2016
1x80mL	Wash buffer Concentrate	RNA43B0	31-Dec-2016
1x90mL	Inkubation buffer	RNB43B0	30-Sep-2016
1x12mL	TMB Substrat	RKN51B0	30-Apr-2017
1x12mL	Stop solution	RKK43B0	31-Dec-2016
1x1mL	Standard Glu-Plasminogen	RBJ51B0	28-Feb-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1	1.07	OK
	Calibrator 5	< 0.2	0.01	OK
Wiederfindung der Kontrollen/ Recovery of controls	Internal control 1	60 – 250 µg/mL	92.5	OK
	Internal control 2	60 – 250 µg/mL	96.0	OK

Durchführung von
Prepared by



07-Jul-2015
Datum / Date:

Freigabe von
Approved by

Fikret NASUFI

07-Jul-2015
Datum / Date: