



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQ...00AZ

PAP Complex ELISA Kit 96T.

REF TC12060
LOT 120601B4.01
 31 – Dec – 2016

Inhalt Content	Reagenz Reagent	LOT	
1 x 12	PAP Plate	RGY33B1	31-Jan-2017
1 x 0.3mL	anti PAP POX	RBK21B2	31-Jan-2017
2 x 20mL	Wash buffer Concentrate	RBE12B2	31-Jan-2017
1 x 20mL	Dilution Buffer	RAF11B2	31-Jan-2017
1 x 12mL	TMB Substrate	RKN33B1	31-Jan-2017
1 x 12mL	Stop solution	RKK32B1	31-Jan-2017
1 x 0,5mL	PAP-Calibrator 1	1BM11B1	31-Jan-2017
1 x 0.5mL	PAP-Calibrator 2	2BM11B1	31-Jan-2017
1 x 0.5mL	PAP-Calibrator 3	3BM11B1	31-Jan-2017
1 x 0.5mL	PAP-Calibrator 4	4BM11B1	31-Jan-2017
1 x 0.5mL	PAP-Calibrator 5	5BM11B1	31-Jan-2017
1 x 0.5mL	PAP high Control	ABM11B1	31-Jan-2017
1 x 0.5mL	PAP low Control	BBM11B1	31-Jan-2017

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result
OD	Calibrator 1	OD: 1 – 2.5 1.85 OK
	Calibrator 5	OD: < 0.2 0.02 OK
Verwendbarkeit Usability	High Control (ABM11B1)	87.9 – 163.1 ng/mL 137.45 OK
	Low Control (BBM11B1)	20.6 – 48.0 ng/mL 33.19 OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

19-Feb-2014
Datum / Date:

Freigabe von
Approved by

Michael RIHA *Michael Riha*

19-FEB-2014
Datum / Date: