



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQ...00AZ

PAP Complex ELISA Kit 96T.

REF TC12060

LOT 120602B5.01

⌚ 31-May-2016

| Inhalt Content | Reagenz Reagent | LOT | ⌚ |
|-------------------|-------------------------|------------|-------------|
| 1 x 12 | PAP Plate | RGY41B0 | 31-May-2016 |
| 1 x 0.3mL | anti PAP POX | RBK41B0 | 31-May-2016 |
| 2 x 20mL | Wash buffer Concentrate | RBE51B0 | 31-Jul-2017 |
| 1 x 20mL | Dilution Buffer | RAF42B0 | 30-Jun-2016 |
| 1 x 12mL | TMB Substrate | RKN51B0 | 30-Apr-2017 |
| 1 x 12mL | Stop solution | RKK51B0 | 30-Jun-2017 |
| 1 x 0,5mL | PAP-Calibrator 1 | 1BM11B1 | 31-Jan-2017 |
| 1 x 0.5mL | PAP-Calibrator 2 | 2BM11B1 | 31-Jan-2017 |
| 1 x 0.5mL | PAP-Calibrator 3 | 3BM11B1 | 31-Jan-2017 |
| 1 x 0.5mL | PAP-Calibrator 4 | 4BM11B1 | 31-Jan-2017 |
| 1 x 0.5mL | PAP-Calibrator 5 | 5BM11B1 | 31-Jan-2017 |
| 1 x 0.5mL | PAP high Control | ABM11B1 | 31-Jan-2017 |
| 1 x 0.5mL | PAP low Control | BBM11B1 | 31-Jan-2017 |

| Analytische Prozedur Analytical Procedure | | Anforderung Requirement | Ergebnis Result | |
|--|------------------------|----------------------------|--------------------|----|
| OD | Calibrator 1 | OD: 1 – 2.5 | 1.32 | OK |
| | Calibrator 5 | OD: < 0.2 | 0.02 | OK |
| Verwendbarkeit Usability | High Control (ABM11B1) | 87.9 – 163.1 ng/mL | 130.7 | OK |
| | Low Control (BBM11B1) | 20.6 – 48.0 ng/mL | 30.8 | OK |

Durchführung von
Prepared by

07-Sep-2015
Datum / Date:

Freigabe von
Approved by

07-Sep-2015
Datum / Date: