



# ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQ...00AZ

## PAP Complex ELISA Kit 96T.

REF      TC12060  
LOT        120603B3.02  
      31 – Mar – 2015

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
1 x 12	PAP Plate	RGY32B0	31-Aug-2015
1 x 0.3mL	anti PAP POX	RBK21B1	30-JUN-2015
2 x 20mL	Wash buffer Concentrate	RBE31B0	30-APR-2015
1 x 20mL	Dilution Buffer	RAF31B0	30-JUN-2016
1 x 12mL	TMB Substrate	RKN32B0	31-MAY-2015
1 x 12mL	Stop solution	RKK32B0	30-Jun-2015
1 x 0,5mL	PAP-Calibrator 1	1BM11B0	31-MAR-2015
1 x 0.5mL	PAP-Calibrator 2	2BM11B0	31-MAR-2015
1 x 0.5mL	PAP-Calibrator 3	3BM11B0	31-MAR-2015
1 x 0.5mL	PAP-Calibrator 4	4BM11B0	31-MAR-2015
1 x 0.5mL	PAP-Calibrator 5	5BM11B0	31-MAR-2015
1 x 0.5mL	PAP high Control	ABM11B0	31-MAR-2015
1 x 0.5mL	PAP low Control	BBM11B0	31-MAR-2015

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result
OD	Calibrator 1	OD: 1 – 2.5      2.2      OK
	Calibrator 5	OD: < 0.2      0.03      OK
Verwendbarkeit Usability	High Control (ABM11B0)	87.9 – 163.1 ng/mL      141.4      OK
	Low Control (BBM11B0)	20.6 – 48.0 ng/mL      35.4      OK

Durchführung von **Fikret NASUFI**  
Prepared by

*F. Nasufi*

15-Oct-2013  
Datum / Date:

Freigabe von  
Approved by

**Michael RIHA**

15-Oct-2013  
Datum / Date: