

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ...00AZ
---	---	-----------

TECHNOZYM® PAP Complex ELISA Kit 96T.

REF TC12060
LOT 120603B5.01
 31-Jan-2017

Inhalt Content	Reagenz Reagent	LOT	
1 x 12	PAP Plate	RGY51B0	31-Jul-2017
1 x 0.3mL	anti PAP POX	RBK41B1	30-Sep-2017
2 x 20mL	Wash buffer Concentrate	RBE51B0	31-Jul-2017
1 x 20mL	Dilution Buffer	RAF51B0	31-Jul-2017
1 x 12mL	TMB Substrate	RKN51B0	30-Apr-2017
1 x 12mL	Stop solution	RKK51B0	30-Jun-2017
1 x 0,5mL	PAP-Calibrator 1	1BM11B1	31-Jan-2017
1 x 0.5mL	PAP-Calibrator 2	2BM11B1	31-Jan-2017
1 x 0.5mL	PAP-Calibrator 3	3BM11B1	31-Jan-2017
1 x 0.5mL	PAP-Calibrator 4	4BM11B1	31-Jan-2017
1 x 0.5mL	PAP-Calibrator 5	5BM11B1	31-Jan-2017
1 x 0.5mL	PAP high Control	ABM11B1	31-Jan-2017
1 x 0.5mL	PAP low Control	BBM11B1	31-Jan-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	OD: 1 – 2.5	1.21	OK
	Calibrator 5	OD: < 0.2	0.02	OK
Verwendbarkeit Usability	High Control (ABM11B1)	87.9 – 163.1 ng/mL	141.68	OK
	Low Control (BBM11B1)	20.6 – 48.0 ng/mL	32.35	OK

Durchführung von
Prepared by

Fikret NASUFI
Fikret NASUFI

14-Oct-2015
Datum / Date:

Freigabe von
Approved by

Michael KAFKA
Michael KAFKA

14-Oct-2015
Datum / Date: