



**ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS**

DQ...00AZ

**PAP Complex ELISA Kit 96T.**

**REF** TC12060

**LOT** 120605B4.01

29-Feb-2016

Inhalt Content	Reagenz Reagent	<b>LOT</b>	
1 x 12	PAP Plate	RGY41B0	31-May-2016
1 x 0.3mL	anti PAP POX	RBK41B0	31-May-2016
2 x 20mL	Wash buffer Concentrate	RBE41B0	29-Feb-2016
1 x 20mL	Dilution Buffer	RAF31B0	30-Jun-2016
1 x 12mL	TMB Substrate	RKN41B0	31-Aug-2016
1 x 12mL	Stop solution	RKK42B0	30-Sep-2016
1 x 0,5mL	PAP-Calibrator 1	1BM11B1	31-Jan-2017
1 x 0.5mL	PAP-Calibrator 2	2BM11B1	31-Jan-2017
1 x 0.5mL	PAP-Calibrator 3	3BM11B1	31-Jan-2017
1 x 0.5mL	PAP-Calibrator 4	4BM11B1	31-Jan-2017
1 x 0.5mL	PAP-Calibrator 5	5BM11B1	31-Jan-2017
1 x 0.5mL	PAP high Control	ABM11B1	31-Jan-2017
1 x 0.5mL	PAP low Control	BBM11B1	31-Jan-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	OD: 1 – 2.5	1.44	OK
	Calibrator 5	OD: < 0.2	0.01	OK
Verwendbarkeit Usability	High Control (ABM11B1)	87.9 – 163.1 ng/mL	138.32	OK
	Low Control (BBM11B1)	20.6 – 48.0 ng/mL	33.11	OK

Durchführung von  
Prepared by

Fikret NASUFI

17-Dec-2014  
Datum / Date:

Freigabe von  
Approved by

Lavdrim NASUFI

17-Dec-2014  
Datum / Date: