

	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ7BM0AZ
---	---	----------

**TECHNOZYM® PAP ELISA Control Set 2x0.5mL**

REF      TC12064  
LOT      7BM41B0.02  
      31-Jan-2017

Inhalt/ Content	Reagenz/Reagent	LOT	
1 x 0.5ml	PAP High Control	ABM11B1	31-Jan-2017
1 x 0.5ml	PAP Low Control	BBM11B1	31-Jan-2017

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Verwendbarkeit / Usability	High Control (ABM11B1)	87.9 – 163.1 ng/mL	138.81    OK
	Low Control (BBM11B1)	20.6 – 48.0 ng/mL	30.34    OK

**Durchführung von**  
**Prepared by**

*Fikret NASUFI*

14-Oct-2015  
Datum / Date:

**Freigabe von**  
**Approved by**

**Michael KAFKA**  
*Michael Kafka*

14-Oct-2015  
Datum / Date: