



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQRD00AZ

PAI-1 Antigen ELISA Kit 96 T.

REF TC12075
LOT RD41B00.01
 31 - Aug -2015

Inhalt Content	Reagenz Reagent	LOT	
1x12mL	KK stop solution for TMB	RKK41B0	31-Mar-2016
1x0,3mL	KL PAI POX For TMB	RKL31B0	31-Aug-2015
1x12mL	KN TMB substrate	RKN34B0	30-Nov-2015
1x80mL	NA wash buffer + thimerosal	RNA41B0	29-Feb-2016
1x90mL	NB Incubation buffer	RNB41B0	31-Jan-2016
1x0.5mL	BQ Calibrator 1 PAI-1	1BQ34B0	31-Oct-2015
1x0.5mL	BQ Calibrator 2 PAI-1	2BQ34B0	31-Oct-2015
1x0.5mL	BQ Calibrator 3 PAI-1	3BQ34B0	31-Oct-2015
1x0.5mL	BQ Calibrator 4 PAI-1	4BQ34B0	31-Oct-2015
1x0.5mL	BQ Calibrator 5 PAI-1	5BQ34B0	31-Oct-2015
1x	GQ-anti PAI-1 coated plate	RGQ41B0	29-Feb-2016
1x0.5mL	ABQ-PAI-1 High Control 0.5 ml	ABQ34B0	31-Oct-2015
1x05mL	BBQ-PAI-1 low Control	BBQ34B0	31-Oct-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	1.0 – 2.5	1.843	OK
	Calibrator 5	< 0.1	0.011	OK
Verwendbarkeit Usability	High Control (ABQ34B0)	61.08 – 101.80 ng/mL	80.136	OK
	Low Control (BBQ34B0)	6.92 – 16.14 ng/mL	11.537	OK

Durchführung von
Prepared by

Michael RIHA

27-Mai-2014
Datum / Date:

Freigabe von
Approved by

Fikret NASUFI

27-Mai-2014
Datum / Date: