

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQRD00AZ
---	---	----------

**PAI-1 Antigen ELISA Kit 96 T.**

REF      TC12075  
LOT      RD51B00.01  
      31-Aug-2016

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
1x	GQ-anti PAI-1 coated plate	RGQ51B0	31-Jan-2017
1x12mL	KK stop solution for TMB	RKK43B0	31-Dec-2016
1x0.3mL	KL PAI POX For TMB	RKL31B1	28-Feb-2017
1x12mL	KN TMB substrate	RKN41B0	31-Aug-2016
1x80mL	NA wash buffer + thimerosal	RNA43B0	31-Dec-2016
1x90mL	NB Incubation buffer	RNB43B0	30-Sep-2016
1x0.5mL	BQ Calibrator 1 PAI-1	1BQ41B0	31-Oct-2016
1x0.5mL	BQ Calibrator 2 PAI-1	2BQ41B0	31-Oct-2016
1x0.5mL	BQ Calibrator 3 PAI-1	3BQ41B0	31-Oct-2016
1x0.5mL	BQ Calibrator 4 PAI-1	4BQ41B0	31-Oct-2016
1x0.5mL	BQ Calibrator 5 PAI-1	5BQ41B0	31-Oct-2016
1x0.5mL	ABQ-PAI-1 High Control	ABQ41B0	31-Oct-2016
1x05mL	BBQ-PAI-1 Low Control	BBQ41B0	31-Oct-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	1.0 – 2.5	2.01	OK
	Calibrator 5	< 0.1	0.04	OK
Verwendbarkeit Usability	High Control (ABQ41B0)	59.04 – 98.40 ng/mL	80.93	OK
	Low Control (BBQ41B0)	6.96 – 16.25 ng/mL	12.68	OK

Durchführung von  
Prepared by

**Fikret NASUFI** *Fikret Nasufi*

25-FEB-2015  
Datum / Date:

Freigabe von  
Approved by

**Michael RIHA** *Michael Riha*

25-FEB-2015  
Datum / Date: