
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQRD00AZ
---	---	----------


TECHNOZYM® PAI-1 ELISA Kit 96T.

REF TC12075
LOT RD52B00.01
 31-Oct-2016

Inhalt Content	Reagenz Reagent	LOT	
1x12mL	KK stop solution for TMB	RKK51B0	30-Jun-2017
1x0,3mL	KL PAI POX For TMB	RKL31B1	28-Feb-2017
1x12mL	KN TMB substrate	RKN51B0	30-Apr-2017
1x80mL	NA wash buffer + thimerosal	RNA51B0	30-Apr-2017
1x90mL	NB Incubation buffer	RNB51B0	28-Feb-2017
1x0.5mL	BQ Calibrator 1 PAI-1	1BQ41B0	31-Oct-2016
1x0.5mL	BQ Calibrator 2 PAI-1	2BQ41B0	31-Oct-2016
1x0.5mL	BQ Calibrator 3 PAI-1	3BQ41B0	31-Oct-2016
1x0.5mL	BQ Calibrator 4 PAI-1	4BQ41B0	31-Oct-2016
1x0.5mL	BQ Calibrator 5 PAI-1	5BQ41B0	31-Oct-2016
1x	GQ-anti PAI-1 coated plate	RGQ52B0	30-Sep-2017
1x0.5mL	ABQ-PAI-1 High Control	ABQ41B0	31-Oct-2016
1x05mL	BBQ-PAI-1 low Control	BBQ41B0	31-Oct-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	1.0 – 2.5	2.03	OK
	Calibrator 5	< 0.1	0.03	OK
Verwendbarkeit Usability	High Control (ABQ41B0)	59.04 – 98.40 ng/mL	82.30	OK
	Low Control (BBQ41B0)	6.96 – 16.25 ng/mL	11.78	OK

Durchführung von
Prepared by



15-Sep-2016

Datum / Date:

Freigabe von
Approved by



15-Sep-2016

Datum / Date: