





**ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS**

DQRXXAZ

TECHNOZYM® t-PA Combi Actibind ELISA Kit 96T

REF TC16000
LOT 160001B6.01
 30-Nov-2016

Inhalt Content	Reagenz Reagent	LOT	
1x	anti t-PA Plate	RGN52B0	30-Sep-2017
1x0.3mL	POX anti t-PA	RKM32B1	30-Nov-2016
1x80mL	Wash buffer Concentrate	RNA52B0	31-Oct-2017
1x90mL	Incubation buffer	RNB61B0	31-Jan-2018
1x12mL	TMB Substrat	RKN52B0	31-Dec-2017
1x12mL	Stop solution	RKK52B0	31-Dec-2017
1x0.5mL	Standard t-PA	RAW11B2	31-Dec-2016
1x20mL	Detection mix buffer- TRIS	RDA51B0	30-Nov-2017
1x	Detection mixture	RBA51B0	31-Dec-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1 activity	> 0.7	1.18	OK
	Calibrator 5 activity	< 0.1	0.05	OK
	Calibrator 1 antigen	> 1.0	1.26	OK
	Calibrator 5 antigen	< 0.1	0.01	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

17-Feb-2016
Datum / Date:

Freigabe von
Approved by

Christoph Enne
Christoph ENNE

17-Feb-2016
Datum / Date: