

TECO medical

TECO® Intact Proinsulin ELISA KIT

Certificate of Analysis

REF
LOT

TE1011
016B



2016-09-30

Refer to Product Insert for further Information

Standard Curve

Calibrators could vary
from batch to batch

Calibrators	Δ pmol/L WHO 09/296	\bar{x} Extinction 450 nm	\bar{x} Extinction 405 nm
A	0.0	0.006	0.009
B	6.6	0.227	0.077
C	11.9	0.406	0.135
D	34	1.169	0.375
E	71	2.416	0.771
F	157	-	1.483
L - Control 1	20,7 (13,5 - 27,9)	0.653	0.215
H - Control 2	47,3 (35,5 - 59,1)	1.554	0.497

ITEM

LOT

EXP

① Specific Intact Proinsulin Antibody coated wells	EBX-DZ-034	2017-01
② Blockingbuffer	EBX-LA-027	2017-04
③ Antibody HRP Conjugate	EBX-AZ-1531	2016-10
④ TMB Substrat	G-ZZ-IB-023	2018-08
⑤ Concentrated wash solution	EBX-HZ-023	2017-04
⑥ Stop Solution	GZZ-KZ-015	2016-09
A Calibrator A	E-BX-B0-017	2017-06
B Calibrator B	E-BX-B1-017	2017-06
C Calibrator C	E-BX-B2-017	2017-06
D Calibrator D	E-BX-B3-017	2017-06
E Calibrator E	E-BX-B4-017	2017-06
F Calibrator F	E-BX-B5-017	2017-06
L Control L	E-BX-BZ-017	2017-06
H Control H	E-BX-BZ-017	2017-06

TECOmedical

Qc release

Date

2015-12-14