

TECOmedical
TECO® ADIPONECTIN ELISA KIT

Certificate of Analysis

REF TE1014
LOT 261015

EXP 2017-11-28

Refer to Product Insert for further Information

Standard Curve

Calibrators	Target Values / Zielwerte / Valeurs attendues	\bar{x} Extinction 450 nm
A	2 ng/ml	0.064
B	10 ng/ml	0.323
C	30 ng/ml	0.921
D	70 ng/ml	1.915
E	100 ng/ml	2.609
L - Control 1	10.37 +/- 2.06 ng/ml x 1:300 dilution = 3.11µg/ml (2.49- 3.73 µg/ml)	
H - Control 2	35.83 +/- 7.17 ng/ml x 1:300 dilution = 10.75 µg/ml (8.6 - 12.9 µg/ml)	

ITEM	LOT	EXP
① Adiponectin Antibody Coated Wells	271015	27-Nov-2018
② Dilution Buffer	011115	1-Dec-2017
③ Antibody HRP Conjugate	291015	29-Nov-2017
④ TMB Substrate	TMBWR09	22-Apr-2019
⑤ Concentrated Wash Solution	111115	11-Nov-2020
⑥ Stop Solution 0,2 M H2SO4	161015	16-Oct-2020
A Calibrators - lyophilised - 2 ng/ml	281015	28-Nov-2017
B Calibrators - lyophilised - 10 ng/ml	281015	28-Nov-2017
C Calibrators - lyophilised - 30 ng/ml	281015	28-Nov-2017
D Calibrators - lyophilised - 70 ng/ml	281015	28-Nov-2017
E Calibrators - lyophilised - 100 ng/ml	281015	28-Nov-2017
L Control 1 - lyophilised	301015	30-Nov-2017
H Control 2 - lyophilised	311015	1-Dec-2017

TECOmedical

Qc release

Date 03.12.2015