

# TECOmedical

## TECO® LEPTIN ELISA KIT

Certificate of Analysis

<b>REF</b>	TE1016	<b>EXP</b>	2016-12-16
<b>LOT</b>	140415/3		

Refer to Product Insert for further Information

### Standard Curve

Calibrators	Target Values / Zielwerte / Valeurs attendues	$\bar{x}$
	ng/mL	Extinction 450 nm
A	1	0.063
B	10	0.599
C	25	1.217
D	50	1.839
E	100	2.497
L - Control 1	11.51 ± 2.30	0.712
H - Control 2	32.5 ± 6,49	1.473

<b>ITEM</b>	<b>LOT</b>	<b>EXP</b>
① Leptin Antibody Coated Wells	140415	14-May-2017
② Dilution Buffer	180415	18-May-2017
③ Antibody HRP Conjugate	190415	19-May-2017
④ TMB Substrate	TMBWP13	18-Jun-2018
⑤ Concentrated Wash Solution	240215	24-Feb-2020
⑥ Stop Solution 0,2 M H2SO4	020315	2-Mar-2020
<b>A</b> Calibrators - lyophilised - 1 ng/ml	161215	16-Jan-2018
<b>B</b> Calibrators - lyophilised - 10 ng/ml	161215	16-Jan-2018
<b>C</b> Calibrators - lyophilised - 25 ng/ml	161215	16-Jan-2018
<b>D</b> Calibrators - lyophilised - 50 ng/ml	161215	16-Jan-2018
<b>E</b> Calibrators - lyophilised - 100 ng/ml	161215	16-Jan-2018
<b>L</b> Control 1 - lyophilised	160415	16-May-2017
<b>H</b> Control 2 - lyophilised	011215	30-Jan-2018

**TECOmedical**

Qc release

21.12.2015