

**Manufactured for Prometheus Laboratories**

The kit includes an extra control (M - Control 3) not mentioned in the general kit instruction.

**TECO medical**

**TECO® HYALURONIC ACID PLUS ELISA KIT**

**For research use only**

Certificate of Analysis

Analysezertifikat

Certificat d'Analyse

**REF**  
**LOT**

TE1018-2

181014

📅 2016-03-31

Refer to Product Insert for further Information

Weitere Informationen hierzu finden Sie in der Packungsbeilage des Produkts

Se référer à la Notice produit pour plus d'informations

Standard Curve

Eichkurve

Courbe Standard

<b>Standards</b>	<b><math>\Delta</math> ng/ml</b>	<b><math>\bar{X}</math> Extinction at 450 nm</b>
A	0	0,183
B	15	0,242
C	50	0,335
D	150	0,699
E	450	1,477
F	1000	2,490
L - Control 1	84,9 (55,2 - 115)	0,464
H - Control 2	380 (285 - 475)	1,327
M - Control 3	198 (148 - 247)	0,870

**ITEM**

**LOT**

**EXP**

❶ HABP Coated Wells	PR071014	2016-04-30
❷ Wash Buffer	WA671014	2016-04-30
❸ Sample Diluent	DN020914	2016-03-31
❹ HABP-HRP Conjugate	EN181014	2016-04-30
❺ TMB Substrate	TA240814	2016-08-24
❻ Stop Solution	SA430814	2016-08-31
A Standard - 0 ng/ml	ZA130914	2016-03-31
B Standard - 15 ng/ml	ZA130914	2016-03-31
C Standard - 50 ng/ml	ZA130914	2016-03-31
D Standard - 150 ng/ml	ZA130914	2016-03-31
E Standard - 450 ng/ml	ZA130914	2016-03-31
F Standard - 1000 ng/ml	ZA130914	2016-03-31
L Control 1	ZA130914	2016-03-31
H Control 2	ZA130914	2016-03-31
M Control 3	ZA130914	2016-03-31

**TECOmedical**

Qc release

Date

2014-11-06