

TECO medical

TECO® HYALURONIC ACID PLUS ELISA KIT

For research use only

Certificate of Analysis

Analysezertifikat

Certificat d'Analyse

REF
LOT

TE1018-2
201114

2016-04-30

Refer to Product Insert for further Information
Weitere Informationen hierzu finden Sie in der Packungsbeilage des Produkts
Se référer à la Notice produit pour plus d'informations

Standard Curve

Eichkurve

Courbe Standard

Standards	Δ ng/ml	\bar{x} Extinction at 450 nm
A	0	0,135
B	15	0,179
C	50	0,275
D	150	0,583
E	450	1,334
F	1000	1,969
L - Control 1	82,4 (54,2 -113)	0,371
H - Control 2	374 (288 - 480)	1,176

ITEM

LOT

EXP

① HABP Coated Wells	PR091114	2016-04-30
② Wash Buffer	WA691114	2016-04-30
③ Sample Diluent	DN041114	2016-04-30
④ HABP-HRP Conjugate	EN201114	2016-04-30
⑦ TMB Substrate	TA271114	2016-08-24
⑧ Stop Solution	SA451114	2016-11-30
A Standard - 0 ng/ml	ZA151114	2016-04-30
B Standard - 15 ng/ml	ZA151114	2016-04-30
C Standard - 50 ng/ml	ZA151114	2016-04-30
D Standard - 150 ng/ml	ZA151114	2016-04-30
E Standard - 450 ng/ml	ZA151114	2016-04-30
F Standard - 1000 ng/ml	ZA151114	2016-04-30
L Control 1	ZA151114	2016-04-30
H Control 2	ZA151114	2016-04-30

TECOmedical

Qc release

Date

2014-11-06