




ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS


DQ8F00AZ

TECHNOTHROMBIN® TGA SUB 5x1.5mL

REF 5006235

LOT 8F51B00.01

 31-Oct-2017

Inhalt Content	Reagenz Reagent	LOT	
5x1.5mL	TGA Substrat	8F51B00	31-Oct-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result		
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1B52B00	TGA RC High LOT 7F51B00	140 - 469	423.6	OK
		TGA RC Low LOT 7D51B00	115 - 359	312.6	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

16-Mar-2016
Datum / Date:

Freigabe von
Approved by

Michael KAFKA *Michael Kafka*

16-MAR-2016
Datum / Date: