
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ2B00AZ
---	---	----------

TECHNOTHROMBIN® TGA Control Low 5x1mL

REF 5006330
LOT 2B61B01.01
 28-Feb-2017

Inhalt Content	Reagenz Reagent	LOT	
5x1mL	TGA Control Low	2B61B01	28-Feb-2017

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	75 - 212	172.2	OK
	42 - 103	71.2	OK

**Durchführung von
Prepared by**

Fikret NASUEI *Fikret Nasuei*

26-Feb-2016
Datum / Date:

**Freigabe von
Approved by**

Michael KAFKA *Michael Kafka*

26-FEB-2016
Datum / Date: