

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0U00AZ.02
---	---	-------------

AK-Calibrant 4x1mL

REF	5010004
LOT	0U41B01.07
	31-Oct-2018

Inhalt Content	Reagenz Reagent	LOT	
1x1mL	AK Calibrant A	1N34B01	31-Oct-2018
1x1mL	AK Calibrant B	3L34B01	31-Oct-2018
1x1mL	AK Calibrant C	3M34B01	31-Oct-2018
1x1mL	AK Calibrant D	3N34B01	31-Oct-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	CCN LOT 1P37B00 Technoplastin HIS %	109.7 – 138.3%	133.1	OK
	CCA LOT 3P38B00 Technoplastin HIS %	25.7 – 32.3%	28.8	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**

Christoph Enne
Christoph ENNE

09. DEZ. 2015

Datum / Date:

**Freigabe von
Approved by**

Alessandro Benigni
Alessandro BENIGNI

09. DEZ. 2015

Datum / Date: