

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ1R00AZ
---	---	----------

**Ceveron® Coagulation Reference, 5x1mL**

REF      5220130  
LOT      1R51B00.02  
      31-Oct-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.0 – 7.8	7.6	OK
Virologie Virology	HbsAg HIV 1,2 Ab HCV	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability ATIII, Chromogenes Substrat: F IIa ATIII, Chromogenic substrate: F IIa	CCN <span style="border: 1px solid black; padding: 1px;">LOT</span> 1P37B00	72.8 – 121.3%	98.3	OK
	CCA <span style="border: 1px solid black; padding: 1px;">LOT</span> 3P38B00	28.5 – 47.5%	38.0	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von  
Prepared by**

  
Lavdrim NASUFI

10-Nov-2015  
Datum / Date:

**Freigabe von  
Approved by**

  
Fikret NASUFI

10-Nov-2015  
Datum / Date: