



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQRB00AZ

TECHNOZYM® vWF:CBA ELISA Kit 96T

REF 5450301

LOT RB63B00.01

30-Jun-2017

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x 12	vWF:CBA Plate Collagen Typ III	RNC62B0	31-Mar-2018
1 x 0.3mL	CBA Conjugate	RNE52B0	30-Jun-2017
1x 80mL	Wash buffer Concentrate	RNA61B0	28-Feb-2018
1 x 90mL	Incubation buffer	RNB61B0	31-Jan-2018
1 x 12mL	TMB Substrate	RKN52B0	31-Dec-2017
1 x 12mL	Stop solution	RKK52B0	31-Dec-2017
1 x 0.5mL	vWF Calibrator 1	1ND31B2	31-Mar-2018
1 x 0.5mL	vWF Calibrator 2	2ND31B2	31-Mar-2018
1 x 0.5mL	vWF Calibrator 3	3ND31B2	31-Mar-2018
1 x 0.5mL	vWF Calibrator 4	4ND31B2	31-Mar-2018
1 x 0.5mL	vWF Calibrator 5	5ND31B2	31-Mar-2018
1 x 0.5mL	vWF High Control	AND31B2	31-Mar-2018
1 x 0.5mL	vWF Low Control	BND31B2	31-Mar-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.21	OK
	Calibrator 5	< 0.1	0.01	OK
Verwendbarkeit Usability	High Control (AND31B2)	0.48 – 0.80 IU/mL	0.70	OK
	Low Control (BND31B2)	0.16 – 0.38 IU/mL	0.27	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

11-Apr-2016
Datum / Date:

Freigabe von
Approved by

Michael KAFKA *Michael Kafka*

11-Apr-2016
Datum / Date: