



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQRP00AZ

TECHNOZYM® ADAMTS-13 INH ELISA Kit 48T

REF 5450451
LOT RP64B0S.01
⌚ 31-May-2017

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x 6	Adamts13 INH Plate	RPS64BS	30-Apr-2018
1 x 0.3mL	Adamts13 INH Conjugate	RPV51B0	31-May-2017
1x 80mL	Wash buffer Concentrate	RNA61B0	28-Feb-2018
1 x 90mL	Incubation buffer	RNB61B0	31-Jan-2018
1 x 12mL	TMB Substrate	RKN52B0	31-Dec-2017
1 x 12mL	Stop solution	RKK52B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Calibrator 1	1PT51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH High Control	RPY51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Low Control	RPZ51B0	31-Dec-2017

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result		
OD	Calibrator 1	> 1.0	1.53	OK
	Calibrator 5	< 0.1	0.01	OK
Verwendbarkeit Usability	High Control (RPY51B0)	54.32 – 81.48 U/mL	70.56	OK
	Low Control (RPZ51B0)	5.27 – 12.29 U/mL	11.68	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

18-Apr-2016
Datum / Date:

Freigabe von
Approved by

Michael KAFKA *Michael Kafka*

18-Apr-2016
Datum / Date: