
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQRR00AZ
---	---	----------

TECHNOZYM® ADAMTS-13 Antigen ELISA Kit 96T.

REF 5450601
LOT RR62B00.01
 30-Nov-2016

Inhalt Content	Reagenz Reagent	LOT	
1 x 12	Adamts13 Ag Plate	RRS51B0	31-Mar-2017
1 x 0.3mL	Adamts13 Ag Conjugate	RQV32B1	30-Nov-2016
1x 80mL	Wash buffer Concentrate	RNA52B0	31-Oct-2017
1 x 90mL	Incubation buffer	RNB61B0	31-Jan-2018
1 x 12mL	TMB Substrate	RKN52B0	31-Dec-2017
1 x 12mL	Stop solution	RKK52B0	31-Dec-2017
1 x 0.5mL	Adamts13 Ag Calibrator 1	1UT53B0	31-Jul-2017
1 x 0.5mL	Adamts13 Ag Calibrator 2	2UT53B0	31-Jul-2017
1 x 0.5mL	Adamts13 Ag Calibrator 3	3UT53B0	31-Jul-2017
1 x 0.5mL	Adamts13 Ag Calibrator 4	4UT53B0	31-Jul-2017
1 x 0.5mL	Adamts13 Ag Calibrator 5	6UT53B0	31-Jul-2017
1 x 0.5mL	Adamts13 Ag High Control	RUY53B0	31-Jul-2017
1 x 0.5mL	Adamts13 Ag Low Control	RUZ53B0	31-Jul-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	0.84	*
	Calibrator 5	< 0.1	0.001	OK
Verwendbarkeit Usability	High Control (RUY53B0)	0.29 – 0.49 IU/mL	0.46	OK
	Low Control (RUZ53B0)	0.05 – 0.11 IU/mL	0.09	OK

*Sonderfreigabe
Durchführung von
Prepared by


Fikret NASUFI 

22-Feb-2016
Datum / Date:

Freigabe von
Approved by


Michael KAFKA 

22-Feb-2016
Datum / Date:

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQRR00AZ
---	---	----------

TECHNOZYM® ADAMTS-13 Antigen ELISA Kit 96T.

REF 5450601
LOT RR62B00.01
 30-Nov-2016

Inhalt Content	Reagenz Reagent	LOT	
1 x 12	Adamts13 Ag Plate	RRS51B0	31-Mar-2017
1 x 0.3mL	Adamts13 Ag Conjugate	RQV32B1	30-Nov-2016
1x 80mL	Wash buffer Concentrate	RNA52B0	31-Oct-2017
1 x 90mL	Incubation buffer	RNB61B0	31-Jan-2018
1 x 12mL	TMB Substrate	RKN52B0	31-Dec-2017
1 x 12mL	Stop solution	RKK52B0	31-Dec-2017
1 x 0.5mL	Adamts13 Ag Calibrator 1	1UT53B0	31-Jul-2017
1 x 0.5mL	Adamts13 Ag Calibrator 2	2UT53B0	31-Jul-2017
1 x 0.5mL	Adamts13 Ag Calibrator 3	3UT53B0	31-Jul-2017
1 x 0.5mL	Adamts13 Ag Calibrator 4	4UT53B0	31-Jul-2017
1 x 0.5mL	Adamts13 Ag Calibrator 5	6UT53B0	31-Jul-2017
1 x 0.5mL	Adamts13 Ag High Control	RUY53B0	31-Jul-2017
1 x 0.5mL	Adamts13 Ag Low Control	RUZ53B0	31-Jul-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	0.84	*
	Calibrator 5	< 0.1	0.001	OK
Verwendbarkeit Usability	High Control (RUY53B0)	0.29 – 0.49 IU/mL	0.46	OK
	Low Control (RUZ53B0)	0.05 – 0.11 IU/mL	0.09	OK

*Sonderfreigabe

Durchführung von
Prepared by

Fikret NASUEI 

22-Feb-2016
Datum / Date:

Freigabe von
Approved by

Michael KAFKA 

22-Feb-2016
Datum / Date: