

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ9600.01
---	---	-----------

Ceveron® Cleaning Solution 3% 25 mL

REF 9820300
LOT 9652B00.01
 30-Sep-2016

Analytische Verfahren Analytical Procedure	Anforderung Requirement	Ergebnis Result	
pH	11.5 – 12.5	12.4	OK
Volumen Volume	25.0 – 26.5mL	26.0	OK
Sterilität Sterility	0 CFU/mL	0	OK

Verwendbarkeit, Fibrinogen, Clauss method	CCN LOT 1P37B00	2.52 – 3.78g/L	3.32	OK
Usability, Fibrinogen, Clauss method Ceveron® alpha	CCA LOT 3P38B00	0.93 – 1.39g/L	1.04	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Christoph Enne
Christoph ENNE

19. OKT. 2015
Datum / Date:

Freigabe von
Approved by

Monika Weinbauer
Monika WEINBAUER

19. OKT. 2015
Datum / Date: