

ASSAYS

M30 Apoptosense® ELISA

Prod. No. 10011

M65® ELISA

Prod. No. 10020

M65 EpiDeath® ELISA

Prod. No. 10040

M30 CytoDeath™ ELISA

Prod. No. 10900

ANTIBODIES

M30 CytoDEATH™

- | | |
|----------------|-----------------|
| ▪ Unconjugated | Prod. No. 10700 |
| ▪ Biotin | Prod. No. 10750 |
| ▪ Fluorescein | Prod. No. 10800 |
| ▪ Orange | Prod. No. 10830 |

M5 Keratin 18

Prod. No. 10600


M6 Keratin 18



Prod. No. 10650

Certificate of Analysis**M30 Apoptosense® ELISA****REF 10011**

M30 Apoptosense® ELISA

REF 10011

Kit Lot No. LOT PE-TS 1128	Expiry  2016-10
---	--

Component		A _{450 nm}	LOT	
M30 Standard A	0 U/L	0.104	P-TS 1335	2016-11
M30 Standard B	75 U/L	0.239	P-TS 1336	2016-11
M30 Standard C	150 U/L	0.427	P-TS 1337	2016-11
M30 Standard D	250 U/L	0.702	P-TS 1338	2016-11
M30 Standard E	500 U/L	1.475	P-TS 1339	2016-11
M30 Standard F	750 U/L	2.233	P-TS 1340	2016-11
M30 Standard G	1 000 U/L	2.757	P-TS 1341	2016-11
M5 Coated Microstrips			P-SS 1878	2018-06
M30 Conjugate			P-TS 1352	2016-10
M30 Conjugate Dilution Buffer			P-TS 1331	2016-11
TMB Substrate			P-SS 2272	2018-11
Stop Solution			P-TS 1034	2019-01
Wash Tablet			P-SS 1301	2017-08
Controls	Expected value (U/L)	Obtained value (U/L)	LOT	
M30 Control Low	100 – 150	123	P-TS 1342	2016-11
M30 Control High	550 – 750	666	P-TS 1343	2016-11

Test date: 2016-03-10	Sign: 
Approved by QA: 2016-03-15	Sign: 