




**ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS**

DQRH00AZ

TECHNOZYM® D-Dimer ELISA Kit 96T

REF 2599006
LOT RH63B00.01
 30-Sep-2017

Inhalt Content	Reagenz Reagent	LOT	
1 x	D-Dimer Plate	RKS52B0	30-Sep-2017
1x0.3mL	D-Dimer Conjugate	RKV61B0	31-Mar-2018
1x80mL	Wash buffer Concentrate	RNA61B0	28-Feb-2018
1x90mL	Incubation buffer	RNB61B0	31-Jan-2018
1x12mL	TMB Substrate	RKN52B0	31-Dec-2017
1x12mL	Stop solution	RKK52B0	31-Dec-2017
1x0.5mL	D-Dimer Calibrator 1	1KT31B2	31-Dec-2017
1x0.5mL	D-Dimer Calibrator 2	2KT31B2	31-Dec-2017
1x0.5mL	D-Dimer Calibrator 3	3KT31B2	31-Dec-2017
1x0.5mL	D-Dimer Calibrator 4	4KT31B2	31-Dec-2017
1x0.5mL	D-Dimer Calibrator 5	5KT42B1	31-Dec-2017
1x0.5mL	D-Dimer High Control	RKY31B2	31-Dec-2017
1x0.5mL	D-Dimer Low Control	RKZ31B2	31-Dec-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.02	OK
	Calibrator 5	< 0.1	0.05	OK
Verwendbarkeit Usability	High Control (RKY31B2)	881.4 – 1469.0 ng/mL	1151.9	OK
	Low Control (RKZ31B2)	208.1 – 346.8 ng/mL	311.8	OK

**Durchführung von
Prepared by**

Fikret NASUEI *Fikret Nasuei*

27-Apr-2016
Datum / Date:

**Freigabe von
Approved by**

Michael KAFKA *Michael Kafka*

27-Apr-2016
Datum / Date: