

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ7F00AZ
---	---	----------

TECHNOTHROMBIN® TGA RC High 50x0.5mL

REF 5006216

LOT 7F61B00.02

 28-Feb-2018

Inhalt Content	Reagenz Reagent	LOT	
50x0.5 mL	TGA RC High	7F61B00	28-Feb-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1B52B00	140 - 469	386.3	OK
	TGA Control Low (nM Peak Thrombin) LOT 2B61B01	75 - 212	184.2	OK

**Durchführung von
Prepared by**

Fikret NASUFI *Fikret Nasufi*

26-Feb-2016
Datum / Date:

**Freigabe von
Approved by**

Michael KAFKA *Michael Kafka*

26. Feb - 2016
Datum / Date: