



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ7G00AZ

TECHNOTHROMBIN<sup>®</sup> TGA RD 50x1.5mL

REF 5006222  
LOT 7G51B00.02  
⌚ 31-Jan-2017

Inhalt Content	Reagenz Reagent	LOT	⌚
50x1.5mL	TGA Reagent RD	7G51B00	31-Jan-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1B52B00	341 – 481	472.6	OK
	TGA Control Low (nM Peak Thrombin) LOT 2B61B01	228 - 347	274.8	OK

Durchführung von  
Prepared by

Fikret NASUFI *Fikret Nasufi*

03-May-2016  
Datum / Date:

Freigabe von  
Approved by

Michael KAFKA *Michael Kafka*

03-MAY-2016  
Datum / Date: