



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ7D00AZ

Ceveron[®] TGA RC Low 5x1 mL

REF 5006224

LOT 7D61B.01

⌚ 31-Mar-2018

Inhalt Content	Reagenz Reagent	LOT	⌚
5x1 mL	Ceveron [®] TGA RC Low	7D61B	31-Mar-2018

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1B52B00	92 - 177 150.6 OK
	TGA Control High Lag time (min) LOT 1B52B00	3.8 - 6.0 5.3 OK
	TGA Control High (AUC) LOT 1B52B00	1633 - 2500 2097.8 OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

25-Apr-2016
Datum / Date:

Freigabe von
Approved by

Christoph ENNE *Christoph Enne*

25. APR. 2016
Datum / Date: