



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ8F00AZ

TECHNOTHROMBIN® TGA SUB 50x1.5mL

REF 5006230
LOT 8F52B00.02
⌚ 31-Oct-2017

Inhalt Content	Reagenz Reagent	LOT	⌚
50x1.5mL	TGA Substrat	8F52B00	31-Oct-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1B52B00	TGA RC High LOT 7F51B00 140 - 469	442.8	OK
		TGA RC Low LOT 7D51B00 115 - 359	349.3	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

27-Apr-2016
Datum / Date:

Freigabe von
Approved by

Michael KAFKA *Michael Kafka*

21-Apr-2016
Datum / Date: