



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ8F00AZ

TECHNOTHROMBIN<sup>®</sup> TGA SUB 5x1.5mL

REF 5006235  
LOT 8F52B00.01  
⌚ 31-Oct-2017

Inhalt Content	Reagenz Reagent	LOT	⌚
5x1.5mL	TGA Substrat	8F52B00	31-Oct-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1B52B00	TGA RC High LOT 7F51B00 140 - 469	397.9	OK
		TGA RD Reagenz LOT 7G51B00 341 - 481	472.6	OK

Durchführung von  
Prepared by

Fikret NASUFI *Fikret Nasufi*

02-May-2016  
Datum / Date:

Freigabe von  
Approved by

Michael KAFKA *Michael Kafka*

03-MAY-2016  
Datum / Date: