
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ2B00AZ
---	---	----------

TECHNOTHROMBIN[®] TGA Control Low 5x1mL

REF 5006330

LOT 2B61B01.02

 28-Feb-2017

Inhalt Content	Reagenz Reagent	LOT	
5x1mL	TGA Control Low	2B61B01	28-Feb-2017


Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Reagenz RC High (nm Peak Thrombin) LOT 7F51B00	75 - 212	101.1 OK
	TGA Reagenz RC Low (nm Peak Thrombin) LOT 7D51B00	42 - 103	71.9 OK

Durchführung von
Prepared by


Lucas ROSENMAYR

22. APR. 2016
 Datum / Date:

Freigabe von
Approved by


Fikret NASUFI

22. APR. 2016
 Datum / Date: