

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ4800AZ.01
---	---	-------------

F VIII Mangelplasma nativ, 50x1mL

REF 5154016
LOT 4861B00.01
 31-Jan-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.8	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Positive	Negative Negative Positive	OK
Verwendbarkeit Usability Ceveron® alpha	CCN LOT 1P37B00	65.3 – 108.8 %	82.3	OK
	CCA LOT 3P38B00	15.0 – 25.0 %	24.0	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**


Monika WEINBAUER

29. FEB. 2016

Datum / Date:

**Freigabe von
Approved by**


Lucas ROSENMAYR

29. FEB 2016

Datum / Date: