



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ1R00.01

Ceveron Coagulation Reference, 5x1mL

REF 5220110

LOT 1R51B01.01

28-Feb-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.0 – 7.8	7.6	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron® alpha	Fibrinogen, Clauss meth. CCN [LOT] 1P37B00	2.52 – 3.78 g/L	3.24	OK
	Fibrinogen, Clauss meth. CCA [LOT] 3P38B00	0.93 – 1.39 g/L	1.22	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Lavdrin NASUFI

18. APR. 2016

Datum / Date:

Freigabe von
Approved by

Lucas ROSENMAYR

18. APR. 2016

Datum / Date: