

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ0Z00AZ.02
---	---	-------------

## TECHNOCHROM<sup>®</sup> FXIII 3 x 3mL

REF            5360010  
LOT            0Z61B00.01  
            28-Feb-2017

Inhalt	Reagenz	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
3 x 3 ml	FXIII Activator Reagent	1473100/1	28-Feb-2017
3 x 3 ml	FXIII Detection Reagent	1473100/2	28-Feb-2017
3 x 3 ml	FXIII NADPH	1473100/3	28-Feb-2017
3 x 1 ml	FXIII Inhibitor Reagent	1473100/4	28-Feb-2017
1 x 6 ml	Stabilizer solution	1473100/5	28-Feb-2017

Test	Spezifikation	Ergebnis
Wiederfindung der Kontrollen / Recovery of controls	Coagulation Control N Lot: 1P37B00	60.8% – 101.3%            67.4%
	Coagulation Control A Lot: 3P38B00	28.5% – 47.5%            39.2%

Durchführung von  
Prepared by

Alessandro BENIGNI

11. FEB. 2016

Datum / Date:

Freigabe von  
Approved by

Monika WEINBAUER

11. Feb. 2016  
Datum / Date:



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ7D00AZ

TECHNOTHROMBIN® TGA RC Low 5x0.5mL

REF 5006212

LOT 7D51B00.02

⌚ 28-Feb-2017

Inhalt Content	Reagenz Reagent	LOT	⌚
5x0.5 mL	TGA RC Low	7D51B00	28-Feb-2017

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1B52B00	115 - 359 302.7 OK

Durchführung von  
Prepared by

Michael KAFKA

23. MAR. 2016  
Datum / Date:

Freigabe von  
Approved by

Christoph ENNE

24. MRZ. 2016  
Datum / Date: