

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0Z00AZ.02
---	---	-------------

TECHNOCHROM[®] FXIII 3 x 3mL

REF 5360010
LOT 0Z63B00.01
 30-Apr-2017

Inhalt	Reagenz	LOT	
3 x 3 mL	FXIII Activator Reagent	KTM1535100	30-Apr-2017
3 x 3 mL	FXIII Detection Reagent	KTM1535100	30-Apr-2017
3 x 3 mL	FXIII NADPH	KTM1535100	30-Apr-2017
3 x 1 mL	FXIII Inhibitor Reagent	KTM1535100	30-Apr-2017
1 x 6 mL	Stabilizer solution	KTM1535100	30-Apr-2017

Test	Spezifikation	Ergebnis
Wiederfindung der Kontrollen / Recovery of controls	Coagulation Control N Lot: 1P37B00	60.8% – 101.3%
	Coagulation Control A Lot: 3P38B00	28.5% – 47.5%

Durchführung von
Prepared by

Michael KAFKA 

28 APR 2016
Datum / Date:

Freigabe von
Approved by

Christoph ENNE 

28. APR. 2016
Datum / Date: