



	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQRU00AZ
---	---	----------

TECHNOZYM® ADAMTS-13 Activity ELISA 96T.

REF 5450701
LOT RU64B00.01
 31-Jul-2017

Inhalt/ Content	Reagenz/Reagent	LOT	
1 x 12	Adamts13 Activity Strips	615	31-Oct-2017
1 x 12mL	Adamts13 Activity HRP Conj antibody	615	31-Oct-2017
1x 53mL	Wash buffer concentrate	615	31-Oct-2017
1 x 30mL	Reaction buffer	615	31-Oct-2017
1 x 12mL	Color Reagent	615	31-Oct-2017
1 x 12mL	Stop solution	615	31-Oct-2017
2 x 6mL	ADAMTS-13 Activity vWF73 Substrate	615	31-Oct-2017
1 x	Sample Dilution MTP	615	31-Oct-2017
1 x 0.5mL	Adamts13 Activity Calibrator 1	1UT51B0	31-Jul-2017
1 x 0.5mL	Adamts13 Activity Calibrator 2	2UT51B0	31-Jul-2017
1 x 0.5mL	Adamts13 Activity Calibrator 3	3UT51B0	31-Jul-2017
1 x 0.5mL	Adamts13 Activity Calibrator 4	4UT51B0	31-Jul-2017
1 x 0.5mL	Adamts13 Activity Calibrator 5	5UT51B0	31-Jul-2017
1 x 0.5mL	Adamts13 Activity Calibrator 6	6UT51B0	31-Jul-2017
1 x 0.5mL	Adamts13 Activity High Control	RUY51B0	31-Jul-2017
1 x 0.5mL	Adamts13 Activity Low Control	RUZ51B0	31-Jul-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	Top OD Cal 1	> 1	1.00	OK
	OD Cal 6	< 0.2	0.02	OK
	High Control (RUY51B0)	0.55 – 0.91 IU/mL	0.58	OK
	Low Control (RUZ51B0)	0.13 – 0.29 IU/mL	0.22	OK

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQRU00AZ
---	---	----------

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

17-May-2016
Datum / Date:

Freigabe von
Approved by

Christoph Enne
Christoph ENNE

17. MAI 2016
Datum / Date: