

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ7UT0AZ.02
---	---	-------------

TECHNOZYM[®] ADAMTS-13 Activity Control Set 2x0.5mL

REF 5450763

LOT 7UT52B0.03

 31-Jul-2017

Inhalt Content	Reagenz Reagent	LOT	
1x0.5mL	Adamts-13 Activity High Control	RUY51B0	31-Jul-2017
1x0.5mL	Adamts-13 Activity Low Control	RUZ51B0	31-Jul-2017

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Verwendbarkeit / Usability	High Control (RUY51B0)	0.55 – 0.91 IU/mL	0.61 OK
	Low Control (RUZ51B0)	0.13 – 0.29 IU/mL	0.22 OK

**Durchführung von
Prepared by**

Fikret NASUFI *Fikret Nasufi*

17-May-2016
Datum / Date:

**Freigabe von
Approved by**

Christoph Enne
Christoph ENNE

17. MAI 2016
Datum / Date: