
	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ3P00AZ.03
---	---	-------------

### Coagulation Control A, 5x1mL

REF      5021055  
LOT      3P38B00.04  
      30-Sep-2017

<b>Analytische Prozedur Analytical Procedure</b>		<b>Anforderung Requirement</b>	<b>Ergebnis Result</b>	
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron® alpha	Siron LS, sec	62.3 – 76.9	66.6	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von  
Prepared by**

  
**Lavdrim NASUFI**

02. MAI 2016  
Datum / Date:

**Freigabe von  
Approved by**

  
**Christoph ENNE**

02. MAI 2016  
Datum / Date: