



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ1000.01

TECHNOVIEW Apixaban CAL 5x1mL

REF 5090269

LOT 1062B00.01

28-Feb-2018

Inhalt Content	Reagenz Reagent	LOT	
1x1mL	TECHNOVIEW Apixaban Cal 1	1161B00	28-Feb-2018
1x1mL	TECHNOVIEW Apixaban Cal 2	1261B00	31-Mar-2018
1x1mL	TECHNOVIEW Apixaban Cal 3	1361B00	30-Apr-2018
1x1mL	TECHNOVIEW Apixaban Cal 4	1461B00	28-Feb-2018
1x1mL	TECHNOVIEW Apixaban Cal 5	1561B00	31-Mar-2018

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	TECHNOVIEW Apixaban Cont low [LOT]3943B01	114.1 – 211.9 ng/mL	163.8 OK
	TECHNOVIEW Apixaban Cont high [LOT]3D41B01	261.2 – 435.4 ng/mL	340.2 OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK	

Durchführung von
Prepared by

Christoph ENNE

09. MAI 2016

Datum / Date:

Freigabe von
Approved by

Monika WEINBAUER

09. MAI 2016

Datum / Date: