

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3D00.01
---	---	-----------

TECHNOVIEW® Apixaban CONT High 5x1mL

REF 5090270
LOT 3D61B00.01
 30-Apr-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.8	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		1.69 – 2.81 IU/mL	2.25	OK
		201.2 – 335.4 ng/mL	268.3	
Verwendbarkeit Usability Ceveron ® alpha	Anti-Xa Apixaban	201.2 – 335.4 ng/mL	310.2	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Lucas ROSENMAYR

20. JUNI 2016

Datum / Date:

Freigabe von
Approved by


Christoph ENNE

20. JUNI 2016

Datum / Date: