


|                                                                                   |                                                       |           |
|-----------------------------------------------------------------------------------|-------------------------------------------------------|-----------|
|  | <b>ANALYSENZERTIFIKAT<br/>CERTIFICATE OF ANALYSIS</b> | DQ3900.01 |
|-----------------------------------------------------------------------------------|-------------------------------------------------------|-----------|

**TECHNOVIEW® Apixaban CONT Low 5x1ml**

REF      5090271  
LOT      3961B00.01  
      30-Apr-2018

| Analytische Prozedur<br>Analytical Procedure                            |                                                     | Anforderung<br>Requirement       | Ergebnis<br>Result               |    |
|-------------------------------------------------------------------------|-----------------------------------------------------|----------------------------------|----------------------------------|----|
| pH                                                                      |                                                     | 7.4 – 7.8                        | 7.8                              | OK |
| Virologie<br>Virology                                                   | HBsAg / HBsAg<br>HIV-AK / HIV-Ab<br>HCV-AK / HCV-Ab | Negative<br>Negative<br>Negative | Negative<br>Negative<br>Negative | OK |
| Chargenspezifische Kalibrationswerte<br>Lot specific calibration values |                                                     | 0.57 – 1.05 U/mL                 | 0.81                             | OK |
|                                                                         |                                                     | 74.1 – 137.5 ng/mL               | 105.8                            |    |
| Verwendbarkeit<br>Usability<br>Ceveron ® alpha                          | Anti-Xa Apixaban                                    | 74.1 – 137.5 ng/mL               | 119.0                            | OK |
| Etiketten und Barcode Überprüfung<br>Label and bar code Verification    |                                                     |                                  | OK                               |    |

**Durchführung von  
Prepared by**

*Lucas Rosenmayr*  
**Lucas ROSENMAYR**

20. JUNI 2016

Datum / Date:

**Freigabe von  
Approved by**

*Christoph Enne*  
**Christoph ENNE**

20. JUNI 2016

Datum / Date: