




ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS


DQ0J00AZ.01

TECHNOCHROM[®] C1-Inhibitor 30T.

REF 5345003

LOT 0J63B00.02

 30-Sep-2017

Inhalt Content	Reagenz Reagent	LOT	
1x1mL	Coagulation Control N	1P37B00	30-Sep-2017
1x1mL	Coagulation Control A	3P38B00	30-Sep-2017
1x0.825mL	Coagulation Reference	1R51B01	28-Feb-2018
1x3mL	C1 Esterase	7B63B00	31-May-2018
1x3mL	Substrate f. C1-esterase	8C61B00	30-Apr-2021
1x25mL	C1-Inh Probenpuffer A	9M51B00	30-Jun-2020
1x20mL	C1-Inh Reaktionspuffer B	9N51B00	31-Jul-2020

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron [®] alpha	CCN LOT 1P37B00	77.3 – 128.8%	104.1	OK
	CCA LOT 3P38B00	29.3 – 48.8%	41.6	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	


Durchführung von
Prepared by


Christoph ENNE

20. JUNI 2016

Datum / Date:

Freigabe von
Approved by


Lucas ROSENMAYR

20. JUNI 2016

Datum / Date: