




**ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS**

DQRP00AZ

**TECHNOZYM® ADAMTS-13 INH ELISA Kit 96T**

**REF** 5450401  
**LOT** RP66B00.01  
 31-May-2017

Inhalt Content	Reagenz Reagent	<b>LOT</b>	
1 x 12	Adamts13 INH Plate	RPS65B0	30-Apr-2018
1 x 0.3mL	Adamts13 INH Conjugate	RPV51B0	31-May-2017
1x 80mL	Wash buffer Concentrate	RNA61B0	28-Feb-2018
1 x 90mL	Incubation buffer	RNB61B0	31-Jan-2018
1 x 12mL	TMB Substrate	RKN52B0	31-Dec-2017
1 x 12mL	Stop solution	RKK52B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Calibrator 1	1PT51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH High Control	RPY51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Low Control	RPZ51B0	31-Dec-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.42	OK
	Calibrator 5	< 0.1	0.02	OK
Verwendbarkeit Usability	High Control (RPY51B0)	54.32 – 81.48 U/mL	68.54	OK
	Low Control (RPZ51B0)	5.27 – 12.29 U/mL	10.62	OK

**Durchführung von  
Prepared by**

Fikret NASUEI *Fikret Nasuei*

04-May-2016  
Datum / Date:

**Freigabe von  
Approved by**

Michael KAFKA *Michael Kafka*

04-MAY-2016  
Datum / Date: