



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQRU00AZ

TECHNOZYM® ADAMTS-13 Activity ELISA 96T.

REF 5450701
 LOT RU68B00.01
 ⏳ 30-Nov-2017

Inhalt/ Content	Reagenz/Reagent	LOT	⏳
1 x 12	Adamts13 Activity Strips	620	30-Nov-2017
1 x 12mL	Adamts13 Activity HRP Conj antibody	620	30-Nov-2017
1x 53mL	Wash buffer concentrate	620	30-Nov-2017
1 x 30mL	Reaction buffer	620	30-Nov-2017
1 x 12mL	Color Reagent	620	30-Nov-2017
1 x 12mL	Stop solution	620	30-Nov-2017
2 x 6mL	ADAMTS-13 Activity vWF73 Substrate	620	30-Nov-2017
1 x	Sample Dilution MTP	620	30-Nov-2017
1 x 0.5mL	Adamts13 Activity Calibrator 1	1UT61B0	31-May-2018
1 x 0.5mL	Adamts13 Activity Calibrator 2	2UT61B0	31-May-2018
1 x 0.5mL	Adamts13 Activity Calibrator 3	3UT61B0	31-May-2018
1 x 0.5mL	Adamts13 Activity Calibrator 4	4UT62B0	31-May-2018
1 x 0.5mL	Adamts13 Activity Calibrator 5	5UT61B0	31-May-2018
1 x 0.5mL	Adamts13 Activity Calibrator 6	6UT61B0	31-May-2018
1 x 0.5mL	Adamts13 Activity High Control	RUY61B0	31-May-2018
1 x 0.5mL	Adamts13 Activity Low Control	RUZ61B0	31-May-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	Top OD Cal 1	> 1	1.41	OK
	OD Cal 6	< 0.2	0.02	OK
	High Control (RUY61B0)	0.62 – 1.03 IU/mL	0.91	OK
	Low Control (RUZ61B0)	0.17 – 0.41 IU/mL	0.33	OK



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQRU00AZ

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

21-Jun-2016
Datum / Date:

Freigabe von
Approved by

Michael KAFKA *Michael Kafka*

21-Jun-2016
Datum / Date: