
	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ9600.01
---	---	-----------

**Ceveron® Cleaning Solution 3% 25 mL**

REF      9820300  
LOT      9661B00.01  
      30-Apr-2017

Analytische Verfahren Analytical Procedure	Anforderung Requirement	Ergebnis Result	
pH	11.5 – 12.5	12.4	OK
Volumen Volume	25.0 – 26.5mL	25.0	OK
Sterilität Sterility	0 CFU/mL	0	OK

Verwendbarkeit, Technoclot DTI	TECHNOVIEW Dabigatran Cont low <span style="border: 1px solid black; padding: 1px;">LOT</span> 3745B00	116.0 – 215.4 ng/mL	174.0	OK
Usability, Technoclot DTI Ceveron® alpha	TECHNOVIEW Dabigatran Cont high <span style="border: 1px solid black; padding: 1px;">LOT</span> 3642B01	236.7 – 394.5 ng/mL	337.7	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von  
Prepared by

  
**Lucas ROSENMAYR**

02. JUNI 2016

Datum / Date:

Freigabe von  
Approved by

  
**Alessandro BENIGNI**

02. JUNI 2016

Datum / Date: