

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ7D00AZ
---	---	----------

TECHNOTHROMBIN® TGA RC Low 5x0.5mL

REF 5006212
LOT 7D62B00.01
 31-May-2018

Inhalt Content	Reagenz Reagent	LOT	
5x0.5 mL	TGA RC Low	7D62B00	31-May-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1B52B00	115 - 359	316.1	OK
	TGA Control Low (nM Peak Thrombin) LOT 2B61B02	42 - 103	85.9	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

01-Jul-2016
Datum / Date:

Freigabe von
Approved by

Michael KAFKA *Michael Kafka*

01-Jul-2016
Datum / Date: